

Report of Analysis

Client Sample ID: FIELD BLANK-HCJ	Date Sampled: 04/05/17
Lab Sample ID: JC40473-4	Date Received: 04/05/17
Matrix: DW - Drinking Water FB	Percent Solids: n/a
Project: Jackson BOE Round 2 Jackson, NJ	

44
4

Total Metals Analysis

Analyte	Result	MCL	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Lead	< 0.00050	0.015	0.00050	mg/l	1	04/08/17	04/08/17 MA	EPA 200.8 ¹	EPA 200.8 ²

(1) Instrument QC Batch: MA41752

(2) Prep QC Batch: MP99781

RL = Reporting Limit
MCL = Maximum Contamination Level (NJAC 7:10 11/04)

Report of Analysis

Client Sample ID: HCJ-POE	Date Sampled: 04/05/17
Lab Sample ID: JC40473-5	Date Received: 04/05/17
Matrix: DW - Drinking Water	Percent Solids: n/a
Project: Jackson BOE Round 2 Jackson, NJ	

4.5
4

Total Metals Analysis

Analyte	Result	MCL	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Lead	0.0136	0.015	0.00050	mg/l	1	04/08/17	04/08/17 MA	EPA 200.8 ¹	EPA 200.8 ²

(1) Instrument QC Batch: MA41752

(2) Prep QC Batch: MP99781

RL = Reporting Limit
MCL = Maximum Contamination Level (NJAC 7:10 11/04)

Report of Analysis

Client Sample ID: HCJ-POE FLUSH	Date Sampled: 04/05/17
Lab Sample ID: JC40473-6	Date Received: 04/05/17
Matrix: DW - Drinking Water	Percent Solids: n/a
Project: Jackson BOE Round 2 Jackson, NJ	

4.6
4

Total Metals Analysis

Analyte	Result	MCL	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Lead	0.00107	0.015	0.00050	mg/l	1	04/11/17	04/11/17 MS	EPA 200.8 ¹	EPA 200.8 ²

(1) Instrument QC Batch: MA41770

(2) Prep QC Batch: MP99850

RL = Reporting Limit
MCL = Maximum Contamination Level (NJAC 7:10 11/04)

Report of Analysis

Client Sample ID: HCJ-WF-15	Date Sampled: 04/05/17
Lab Sample ID: JC40473-7	Date Received: 04/05/17
Matrix: DW - Drinking Water	Percent Solids: n/a
Project: Jackson BOE Round 2 Jackson, NJ	

4.7
4

Total Metals Analysis

Analyte	Result	MCL	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Lead	0.00612	0.015	0.00050	mg/l	1	04/08/17	04/08/17 MA	EPA 200.8 ¹	EPA 200.8 ²

(1) Instrument QC Batch: MA41752

(2) Prep QC Batch: MP99781

RL = Reporting Limit
MCL = Maximum Contamination Level (NJAC 7:10 11/04)

Report of Analysis

Client Sample ID: HCJ-WF-15 FLUSH	Date Sampled: 04/05/17
Lab Sample ID: JC40473-8	Date Received: 04/05/17
Matrix: DW - Drinking Water	Percent Solids: n/a
Project: Jackson BOE Round 2 Jackson, NJ	

4.8
4

Total Metals Analysis

Analyte	Result	MCL	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Lead	< 0.00050	0.015	0.00050	mg/l	1	04/08/17	04/08/17 MA	EPA 200.8 ¹	EPA 200.8 ²

(1) Instrument QC Batch: MA41752

(2) Prep QC Batch: MP99781

RL = Reporting Limit
MCL = Maximum Contamination Level (NJAC 7:10 11/04)